The Silicon Vertex Detector of the Belle II Experiment

Yuma Uematsu (UTokyo) on behalf of the Belle II SVD group

uematsu@hep.phys.s.u-tokyo.ac.jp

Outline

Introduction


Operation & Performance

Beam background & Radiation effect

Conclusions

Introduction

Belle II/SuperKEKB




Belle II Vertex Detector

Requirements

 Better vertex resolution than Belle to compensate reduced Lorentz boost
 improved point resolution/reduced inner radius/lower material
 Operate in high background environment
 Hit rate: 3 MHz/cm² @ SVD layer-3
 Radiation hard
 0.2 Mrad/yr. @ SVD layer-3

- Layers-1-2: Pixel Detector (PXD)
 DEPFET pixel sensors
 Innermost layer 1.4 cm from interaction point
- Layers-3-6: Silicon Vertex Detector (SVD)

Diamond sensors
 For radiation monitor and beam abort



935 mm

-15 -10 -5 0 5 10 15 [cm] ←

-5 -10

O Diamond radiation sensor

Belle II Silicon Vertex Detector (SVD)


Layer-3-6: Silicon Vertex Detector (SVD)
 Double-sided Si strip detectors (DSSDs)
 Low material budget: 0.7% X₀ per layer

SVD Roles


- Extrapolate tracks to PXD
 - essential for reconstruction of decay vertices
 PXD region of interest for data reduction
- Stand-alone tracking for low p_T tracks
- Precise vertexing of K_s
- PID with d*E*/dx


DSSDs

- Provide 2-D spatial information
- Strips are AC-coupled to n-type substrate
- Fully depleted at 20-60 V, operated at 100 V
- Total: 172 sensors = 1.2 m² sensor area = 224k readout strips









readout Al

Belle II Silicon Vertex Detector (SVD)


Layer-3-6: Silicon Vertex Detector (SVD)
 Double-sided Si strip detectors (DSSDs)
 Low material budget: 0.7% X₀ per layer

SVD Roles

- Extrapolate tracks to PXD
 - essential for reconstruction of decay vertices
 - PXD region of interest for data reduction
- Stand-alone tracking for low p_T tracks
- Precise vertexing of K_s
- PID with d*E*/dx

DSSDs

- Provide 2-D spatial information
- Strips are AC-coupled to n-type substrate
- Fully depleted at 20-60 V, operated at 100 V
- Total: 172 sensors = 1.2 m² sensor area = 224k readout strips




	125	125	126 [mm]
172 sensors	40	60	61 <mark>←→</mark> 41
	x14	x120	x38
	Small	Large	Trapezoidal
# of p-strips*	768	768	768
p-strip pitch*	50 µm	75 µm	50-75 μm
# of n-strips*	768	512	512
n-strip pitch*	160 µm	240 µm	240 µm
thickness	320 µm	320 µm	300 µm
manufacturer	HPK		Micron

*readout strips - one floating strip on both sides


Front-end ASIC: APV25

- Originally developed for CMS silicon tracker
- Fast: 50 ns shaping times
- Radiation hard: > 100 Mrad
- Power consumption: 0.4 W/chip (700 W in total)
- 128 channel inputs per chip
- Operated in "multi-peak" mode @ ~32 MHz
 - Bunch-crossing frequency ~8*32 MHz, clock not synchronous with them as in CMS
 - 6 subsequent samples recorded (3-sample also possible)
- 3/6-mixed acquisition mode prepared for higher luminosity
 Switching sampling number according to the timing precision of trigger
 Reduce background occupancy/trigger dead-time/data-size
 Half time-window/FIFO usage/hit-information in 3-sample
 - The functionality already implemented in the real setup and confirmed to work
 - Performance study needed before moving to 3/6-mixed mode
 - Hit efficiency is the first step \rightarrow P16
 - To be checked: position resolution, dE/dx, hit-time



Front-end ASIC: APV25

- Originally developed for CMS silicon tracker
- Fast: 50 ns shaping times
- Radiation hard: > 100 Mrad
- Power consumption: 0.4 W/chip (700 W in total)
- 128 channel inputs per chip
- Operated in "multi-peak" mode @ ~32 MHz
 - Bunch-crossing frequency ~8*32 MHz, clock not synchronous with them as in CMS
 - 6 subsequent samples recorded (3-sample also possible)
- 3/6-mixed acquisition mode prepared for higher luminosity
 Switching sampling number according to the timing precision of trigger
 Reduce background occupancy/trigger dead-time/data-size
 Half time-window/FIFO usage/hit-information in 3-sample
 - The functionality already implemented in the real setup and confirmed to work
 - Performance study needed before moving to 3/6-mixed mode
 - Hit efficiency is the first step \rightarrow P16
 - To be checked: position resolution, dE/dx, hit-time





Origami chip on sensor concept

Readout chips directly on each middle sensor

- Shorter signal propagation length (smaller capacitance and noise)
- Thinned to 100 μm to reduce material budget
- Wrapped flex to read both sides from the same side

■ Cool only one side with bi-phase –20 °C CO₂





Operation & Performance

Operational experience

SVD installed in 2018, operated since 2019

Reliable and smooth operation without major problems

- Total fraction of masked strips ~ 1%
- One APV25 chip disabled in spring 2019 (out of 1748)
 - \rightarrow fixed by cable reconnection in summer 2019

Excellent detector performance


Hit efficiency stably > 99% in most of the sensors

Reasonable charge loss

u/P side: agrees with MIP considering uncertainty in calibration

•v/N side: 10-30% due to large pitch and floating strip


Very good SNR (most probable value: 13-30)



Cluster position resolution

◆ Preliminary cluster position resolution measured on e⁺e⁻ → µ⁺µ⁻ data
 ■ Estimated from the residual of the cluster position with respect to the track (unbiased)
 ● Effect of the track extrapolation error subtracted


Excellent position resolution in agreement with the expectations from the pitch
 Still room for improvement for the u/P side (work ongoing)



Hit-time resolution

Excellent hit-time resolution with respect to event time
 Event time estimated by central drift chamber (CDC) outside of SVD
 (~ 2.9 ns u/P, ~ 2.4 ns v/N)

Possible to efficiently reject off-time background hits
 Will be used for higher luminosity and background levels




3-sample acquisition mode

Performance

Ideal 3 samples provide enough information as 6 samples

- amplitude peak ADC sample
- hit-time rising edge of the waveform
- Degrades if the trigger timing is largely shifted
 - CDC event time is a good estimator
- 1st step: relative hit efficiency
 (hit efficiency in 3-sample)/(~ in 6-sample)
 Emulate 3-sample mode offline
 Efficiency based on track using CDC, SVD and PXD
 > 99.9% for trigger jitter less than 30 ns
 - almost a whole clock-cycle




3-sample acquisition mode

Performance

Ideal 3 samples provide enough information as 6 samples

- amplitude peak ADC sample
- hit-time rising edge of the waveform
- Degrades if the trigger timing is largely shifted
 - CDC event time is a good estimator
- 1st step: relative hit efficiency (hit efficiency in 3-sample)/(~ in 6-sample)
 - Emulate 3-sample mode offline
 - Efficiency based on track using CDC, SVD and PXD
 - > 99.9% for trigger jitter less than 30 ns
 - almost a whole clock-cycle




3/6

Beam background & Radiation effect

Beam background and hit occupancy

Beam background increases SVD hit occupancy which degrades tracking performance

- Present occupancy limit in layer-3: ~ 3%
 - will be loosened to x~2 using hit-time to reject background
- With current luminosity, average hit occupancy in layer-3 is well under control (< 0.5%)
- Projection of hit occupancy at $L = 8 \times 10^{35} \text{ cm}^{-2} \text{s}^{-1}$ is about 3% in layer-3
 - estimated by scaling MC with data/MC ratio
 - correspond to dose of ~ 0.2 Mrad/smy
 - $^\circ~$ = 1-MeV neutron fluence of $\sim~5{\times}10^{11}~n_{eq}/cm^2/smy$
 - smy: Snowmass Year = 10⁷ sec
 - Long term BG extrapolation affected by large uncertainties
 - optimization of collimator settings in MC
 - injection BG not included in data nor MC




Integrated dose

 SVD dose estimated by dose on diamond sensors: 70 krad in layer-3 mid plane (the most exposed to radiation)

Dose estimate based on correlation between SVD occupancy and diamonds dose

Several assumptions and large uncertainty (~ 50%)


1-MeV equivalent neutron fluence:
 1.6×10¹¹ n_{eq}/cm² in first 2.5 years
 assuming dose/neq fluence ratio
 = 2.3×10⁹ n_{eq}/cm²/krad from MC



Int. dose in SVD - New coeff. (from exp. 12 & 16) + EODB correction (from 14/2102)

Radiation effect on leakage current

- Good linear correlation between leakage current and estimated dose
 - Slope: 2-5 µA/cm²/Mrad with large variations due to temperature effects and dose spread among sensors in layer (average dose in layer used in estimate)
 - Same order of magnitude as BaBar measurement (1 μA/cm²/Mrad @ 20 °C)
 - [NIMA 729, 615-701, 2013]
 - Strip noise from leakage current is suppressed by short shaping time (50ns) in APV25
 - comparable to the noise from sensor capacitance even after 10 Mrad irradiation



2021/9/13 - Vertex 2021

Radiation effect on strip noise


Noise increase of 20-25% in layer-3

Not affecting performance

Likely due to radiation effects on sensor surface

• Non-linear increase due to fixed oxide charges that increase inter-strip capacitance, expected to saturate


Saturation seen on v/N side and starting to be seen on u/P



Noise average [e-]

Radiation effect on depletion voltage


v/N side strip noise drops at full depletion
 v/N side insulated only when the n-type bulk is fully depleted
 Over-depletion bias still slightly decrease noise
 by reducing electron accumulation layer on v/N side surface



v/N side

No change in full depletion voltage observed with time
 Consistent with low integrated neutron fluence (~1.6 × 10¹¹ n_{eq}/cm²)

L3.5.1 v/N Side - Strip Noise



L3.5.1 N Side - Noise

2021/9/13 - Vertex 2021

Conclusions

 SVD has been taking data in Belle II since March 2019 smoothly and reliably

Excellent performance in agreement with expectations
 Still some room for improvement in cluster position resolution

 Seen first effects of radiation damage at the expected level but not affecting performance

Ready to cope with increased beam background

Reject off-time background using hit-time

3/6-mixed acquisition mode to reduce dead time, data size and occupancy

Related talks?


- PXD talk by whom?
- Mateuz simulation update (?)
- Robin and Lucas position resolution studies (?)

Backup

Signal charge and signal-to-noise ratio

Signal charge: normalized for the track path length in silicon
 u/P side: agree with MIP considering ~ 15% uncertainty in APV25 gain calibration
 v/N side: 10-30% signal loss due to large pitch and presence of floating strip
 similar in all sensors

SNR: very good in all sensors (most probable value: 13-30)
 u/P side: larger noise due to longer strip length (larger inter-strip capacitance)




2021/9/13 - Vertex 2021



Hit-time estimation

 Neighboring strips over threshold are grouped into 'cluster' APV samples for strips $a_{strip 0}$, $a_{strip 2}$, $a_{strip 2}$ are summed up:

$$a_{\text{cluster }i} = \sum_{\text{strip} \in \text{cluster}} a_{\text{strip }i}$$



3-sample acquisition mode

efficiency

- emulate 3-sample mode offline using trigger timing information
- based on tracks using CDC, SVD and PXD hits
- > 99.4% for trigger jitter less than 30 ns

efficiency compared to 6-sample
 3-sample efficiency/6-sample efficiency
 > 99.9% for trigger jitter less than 30 ns
 almost a whole clock-cycle

